



Ref. Certif. No.

**US-32593-UL**

**IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)  
CB SCHEME**

**CB TEST CERTIFICATE**

Product

Switching Power Supply

Name and address of the applicant

SL POWER ELECTRONICS CORP  
BLDG A  
6050 KING DR  
VENTURA, CA 93003 USA

Name and address of the manufacturer

SL POWER ELECTRONICS CORP  
BLDG A  
6050 KING DR  
VENTURA, CA 93003 USA

Name and address of the factory

Note: When more than one factory, please report on page 2

INDUSTRIAS S L S A DE C V  
CIRCUITO SIGLO XXI 2055  
COL PARQUE INDUSTRIAL EX-XXI  
MEXICALI, BC 21254 Mexico

Additional Information on page 2

Ratings and principal characteristics

See Page 2

Trademark (if any)



Type of Customer's Testing Facility (CTF) Stage used

CTF Stage 3

Model / Type Ref.

SE150AXXYZWW  
See Page 2

Additional information (if necessary may also be reported on page 2)

Class I

Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005

As shown in the Test Report Ref. No. which forms part of this Certificate

E116994-D1020-1/A0/C0-CB issued on 2018-10-18

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2018-10-18

Signature:

Jolanta M. Wroblewska

**Model Details:**

SE150AXXYZWW where S can be M or T, XX represents the output voltage from 12V to 48V, YY represents any number from 00 to 99 or blank, Z represents F, means class I construction, WW represents any number from 00 to 99 or blank.

**Factories:**

SL XIANGHE POWER ELECTRONICS CORP  
No.B-02-03, North side of Landscape Ave, Qibu District,  
Environmental Industrial Park  
Xianghe, 065400 Hebei China

**Ratings:**

Output: See Model Differences for the output ratings

**Model TE150A1251Z01:**

Input: 100-110Vac, 50-60Hz, 2.0A; Output: 12Vdc, 11.25A

Input: 110-240Vac, 50-60Hz, 2.0A; Output: 12Vdc, 12.09A

**Model TE150A1551Z01:**

Input: 100-110Vac, 50-60Hz, 2.0A; Output: 15Vdc, 9.00A

Input: 110-240Vac, 50-60Hz, 2.0A; Output: 15Vdc, 9.67A

**Model TE150A1851Z01:**

Input: 100-110Vac, 50-60Hz, 2.0A; Output: 18Vdc, 7.5A

Input: 110-240Vac, 50-60Hz, 2.0A; Output: 18Vdc, 8.06A

**Model TE150A2451Z01:**

Input: 100-110Vac, 50-60Hz, 2.0A; Output: 24Vdc, 6.05A

Input: 110-240Vac, 50-60Hz, 2.0A; Output: 24Vdc, 6.25A

**Model TE150A4851Z01:**

Input: 100-110Vac, 50-60Hz, 2.0A; Output: 48Vdc, 3.13A

Input: 110-240Vac, 50-60Hz, 2.0A; Output: 48Vdc, 3.13A

**Additional Information:**

Additionally evaluated to EN 60601-1:2006/A1:2013/A12:2014 and IEC 60601-1-6:2010/AMD1:2013.

National Differences specified in the CB Test Report.

**Additional information (if necessary)**

UL (US), 333 Pflingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/names](http://www.ul.com/names)

Date: 2018-10-18

Signature:

Jolanta M. Wroblewska