

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST  
CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE  
CERTIFICATS D'ESSAIS DES EQUIPEMENTS  
ELECTRIQUES (IECEE) METHODE OC

**CB TEST CERTIFICATE****CERTIFICAT D'ESSAI OC**

Product  
Produit

Power Supply

Name and address of the applicant  
Nom et adresse du demandeur

SL POWER ELECTRONICS CORP  
BLDG A, 6050 KING DR  
VENTURA, CA 93003, USA

Name and address of the manufacturer  
Nom et adresse du fabricant

SL POWER ELECTRONICS CORP  
BLDG A, 6050 KING DR  
VENTURA, CA 93003, USA

Name and address of the factory  
Nom et adresse de l'usine

SL XIANGHE POWER ELECTRONICS CORP.  
NO.4, SHUANGXING NORTH ROAD, XIANGHE ECONOMIC  
AND TECHNOLOGICAL DEVELOPMENT ZONE, XIANGHE  
COUNTY, HEBEI PROVINCE  
CHINA

Note: When more than one factory, please report on page 2  
Note: Lorsque il y a plus d'une usine, veuillez utiliser la 2<sup>ème</sup> page

Additional Information on page 2  
See Page 2

Ratings and principal characteristics  
Valeurs nominales et caractéristiques principales

Trademark (if any)  
Marque de fabrique (si elle existe)



SMT

Type of Manufacturer's Testing Laboratories used  
Type de programme du laboratoire d'essais  
constructeur

Model / Type Ref.  
Ref. De type

MINT1250AWWXXYZZ, MINT1250BWWXXYZZ,  
MINT1275AWWXXYZZ, MINT1275BWWXXYZZ  
See Page 2

Additional information (if necessary may also be  
reported on page 2)  
Les informations complémentaires (si nécessaire,,  
peuvent être indiqués sur la 2<sup>ème</sup> page

Additionally evaluated to EN 60601-1:2006/A1:2013; National  
Differences specified in the CB Test Report.

Additional Information on page 2

A sample of the product was tested and found  
to be in conformity with  
Un échantillon de ce produit a été essayé et a été  
considéré conforme à la

IEC 60601-1(ed.3), IEC 60601-1(ed.3);am1

As shown in the Test Report Ref. No. which forms  
part of this Certificate  
Comme indiqué dans le Rapport d'essais numéro de  
référence qui constitue partie de ce Certificat

E116994-4786576038A issued on 2014-10-07

This CB Test Certificate is issued by the National Certification Body  
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2014-10-09

Signature:

Jolanta M. Wroblewska



Ref. Certif. No.

**US-24052-UL**

**Model Details:**

MINT1250AWWXXYZZ, MINT1250BWWXXYZZ, MINT1275AWWXXYZZ, MINT1275BWWXXYZZ Where WW represents the output voltage which may be any number from 12 to 56, XX represents output connector which may be any two alphanumeric digits, Y represents the input connector which may be any letter from A thru Z, ZZ represents non-safety related customer options and RoHS statements which may be any two alphanumeric digits.

**Factories:**

SL XIANGHE POWER ELECTRONICS CORP  
NO.B-02-03,NORTH SIDE OF LANDSCAPE AVENUE, QIBU DISTRICT, ENVIRONMENTAL INDUSTRIAL PARK,  
XIANGHE COUNTY, HEBEI PROVINCE 065400  
CHINA

SL POWER ELECTRONICS MEXICALI , MX  
INDUSTRIAS SL S.A. DE C.V. CALLE: CIRCUITO SIGLO XXI #2055 COLONIA: PARQUE INDUSTRIAL EX-XXI C.P.  
21254 MEXICALI, B.C.  
MEXICO

**Ratings:**

Input: 100-240V~, 50-60Hz, 3.7A  
Output: With 200 LFM (linear feet per minute) forced air flow:  
Main Output: 12Vdc/21.84A to 56Vdc/4.68A, Maximum 262W,  
Fan Output: 12Vdc/1.0A  
Signal: 5VSB/0.2A  
Without forced air flow:  
Main Output: 12Vdc/15A to 56Vdc/3.22A, Maximum 180W,  
Fan Output: 12Vdc/1.0A  
Signal: 5VSB/0.2A

**Additional Information:**

The risk management requirements of the standard were not addressed.

**Additional information (if necessary)  
Information complémentaire (si nécessaire)**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2014-10-09

Signature:

Jolanta M. Wroblewska